


## Introduction



## Methods



## Results

In 2013 (July to December), 42632 containers were sterilized in our establishment (= active containers). In 2014 (July to December), after the storeroom was created, 46955 containers were sterilized.

6221 (14,6%) containers were incompletes in 2013 vs 3004 (6,4%) containers in 2014



Figure 1: Number of incomplete containers sterilized

In 2013, these containers were sterilized 6,7 times before being completed. In 2014, they were sterilized 4,6 times before being completed




Figure 2: Number of sterilization runs before complement

The average number of missing instruments was estimated to :

- 2,8 instruments per container in 2013
- 1,5 instruments per container in 2014



Figure 3: Average number of missing instruments per container

## Discussion

- A satisfaction survey will be send to the surgeons
  - > An interesting way to estimate the level of global improvement
- An inventory will be performed after one year
  - > to assess its financial impact

## Conclusion

- Positive impact on the quality of the containers living the CSU
- This unique IBI is relevant and must be maintained in the CSU